

湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology
Hubei Institute of Measurement and Testing Technology

Hubei Institute of Measurement and Testing Technology

地址 (Add): 湖北省武汉市东湖新技术开发区茅店山中路二号
网址 (Web site): <http://www.hbjl.gov.cn>

邮编 (Post Code): 430223
电话 (Tel): 027-81925136
传真 (Fax): 027-81925137

中国校准服务联合体
China United Calibration Service



CUC

REGISTRATION NO. 002

校准证书 CALIBRATION CERTIFICATE

证书编号: [2020DW03400796]
Certificate No.

委托方名称 Customer	武汉登丰电力科技有限公司
委托方地址 Address	武汉东湖新技术开发区光谷大道78号
器具名称 Name Of Sample	钳形接地电阻测试仪
制造厂商 Manufacturer	武汉登丰电力科技有限公司
型号规格 Model/Type	ETCR2000
器具编号 No Of Sample	20200630

证书专用章
Stamp

湖北省计量测试技术
证书骑缝章



校准日期 Calibration date	2020 年 06 月 29 日 Y M D
建议再校日期 Recommended recal.date	2021 年 06 月 28 日 Y M D

证书批准人 Approved by	耿睿
核验员 Checked by	李梅
校准员 Calibrated by	王强

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



校准数据/结果

Data/Results of Calibration

标准值 (Ω)	显示值 (Ω)	测量不确定度 U_{rel} , $k=2$
0.100	0.097	0.5%
0.200	0.20	
0.500	0.50	
0.800	0.80	
1.00	1.0	
2.00	2.0	
5.00	5.0	
8.00	8.0	
10.0	10.0	
20.0	20.0	
50.0	50.0	
80.0	81.0	
100.0	102	



以下空白